

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: <u>130298</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name: <u>Peter C Steen</u>	ID#: <u>4572</u> Date: <u>6/25/08</u>	
Agency: <u>Palmer Police Department</u> ✓	Phone #: <u>(907) 745-4811</u>	
Instrument Location: <u>Palmer Police Department</u>		
Alco S/N: <u>x172967</u> ✓ Target Value: <u>.076</u> ✓ High Pressure: <u>1100</u>		
B		
Alco Test Value Average:	<u>.081</u> ✓ 1 st Alco	<u>.080</u> ✓ 2 nd Alco
Signature: _____	<u>CSB</u> <u>9/8/08</u>	
(OVER)		

Under the Alaska Rules of Evidence, I certify that (verbally write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Peter C Steen, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

09/08/08
Date

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130298 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

JUNE 25, 2008 ✓ *CSB/108*

OPERATOR'S NAME:

STEEN, PETER, C

OPERATOR'S NUMBER: 4572

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/NI:

VOC

O.L. #: VOC

DEPT/AGENCY: PLM1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .076 ✓

ALCO S/N: X172967 ✓

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 29.88 in		
ALCO TARGET	.075	07:04
BLANK TEST	.000	07:05
INTERNAL STANDARD	VERIFIED	07:05
ALCO TV 29.88 in	.081 ✓	07:06
BLANK TEST	.000	07:06
SUBJECT SAMPLE	.000 <i>CSB</i>	07:07
BLANK TEST	.000 <i>1/10/08</i>	07:08
ALCO TV 29.88 in	.000 ✓	07:09
BLANK TEST	.000	07:09

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

JUNE 25, 2008 ✓

TIME 07:03

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.88 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
PQRSTUVWXYZ{|}~■