

**VERIFICATION OF CALIBRATION REPORT**  
*of DataMaster cdm Breath Test Instrument*  
 State of Alaska  
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 19 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130290</u> ✓
Name <u>Fred J. Whittingham</u>	ID# <u>7286</u> ✓	Date <u>7/8/10</u> ✓
A Agency <u>Whittier Police Dept</u>		Phone # <u>472-2340</u>
Instrument Location <u>Whittier PD</u>		
B Alco S/N <u>X124656</u> ✓		Target Value <u>.081</u> ✓ High Pressure <u>900</u>
Alco Test Values <u>.84</u> ✓		<u>.083</u> ✓
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>		<u>[Signature]</u> 7/19/10

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 7<sup>th</sup> day of July 2010.

[Signature]  
 (Notary Seal Stamp)

Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130298

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130298

JULY 08, 2010

OPERATOR'S NAME: *COB #1910*  
WHITTINGHAM, FRED J  
OPERATOR'S NUMBER: 7286  
SUBJECT'S LAST NAME:  
VOC  
SUBJECT'S FIRST NAME/MI :  
VOC  
O.L. #: VOC  
DEPT/AGENCY: MPD  
CASE/REPORT: VOC  
TEST TYPE: U  
ALCO TARGET VALUE: .081  
ALCO S/N: X124656

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.16 in  
ALCO TARGET .081 15:30  
BLANK TEST .000 15:31  
INTERNAL STANDARD VERIFIED 15:31  
ALCO TV 30.16 in .084 15:32  
BLANK TEST .000 15:32  
SUBJECT SAMPLE *COB #1910* .000 15:33  
BLANK TEST .000 15:34  
ALCO TV 30.16 in .083 15:34  
BLANK TEST .000 15:35

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130298

JULY 08, 2010  
TIME 15:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.16 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./:0123456789:;<=>?@ABCDEFG  
HIJLMOQRSTUVWXYZ\]^\_`abcde fghijklmno  
pqrs tuvwxyz{|}~