

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130298

BMB
9/10/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

SEPTEMBER 01, 2010

OPERATOR'S NAME:
WHITTINGHAM, FRED
OPERATOR'S NUMBER: 7286
SUBJECT'S LAST NAME:
VOC

SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: WPD
CASE/REPORT: VOC
TEST TYPE: 0
ALCO TARGET VALUE: .061
ALCO S-N: K124656

— BREATH ANALYSIS —

.061 ADJUSTED FOR 29.86 in		
ALCO TARGET	.060	18:23
BLANK TEST	.060	18:24
INTERNAL STANDARD	VERIFIED	18:24
ALCO TU 29.86 in	.061	18:25
BLANK TEST	.060	18:25
SUBJECT SAMPLE	.060	18:26
BLANK TEST	.060	18:27
ALCO TU 29.86 in	.062	18:27
BLANK TEST	.060	18:28

BMB
9/10/10

DIAGNOSTIC CHECK



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

SEPTEMBER 01, 2010
TIME 18:20

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/02/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.86 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMPQRSSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■