OCT 0 4 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130298 ✓

Supervisor/Operator Performing the Verification Procedure:

A

Name Fred J. Whittingham ID# 7286 ✓ Date ~~10/11/10~~ 9/30/10 DW 10/4/10

Agency Whittier PD Phone # 472-2340

Instrument Location Whittier PD

B

Alco S/N X124656 ✓ Target Value .081 ✓ High Pressure .899

Alco Test Values .082 ✓ (1st Alco) .082 ✓ (2nd Alco)

Signature [signature]

(OVER)

BMB 10/5/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of DEC, 2010.

[signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

DANIEL L. NICHOLAS Commission # 116780 NOTARY PUBLIC My Commission Expires With Office STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130298 BMB 10/5/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

SEPTEMBER 30, 2010 BMB 10/5/10

OPERATOR'S NAME:
WITTINGHAM-FRED
OPERATOR'S NUMBER: 7286
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: WPD
CASE/REPORT: VOC TEST
TEST TYPE: 0
ALCO TARGET VALUE: .081
ALCO S/N: X124655

--- BREATH ANALYSIS ---

.08 ADJUSTED FOR 29.60 in

ALCO TEST	.080	17:22
BLANK TEST	.000	17:23
INTERNAL STANDARD	VERIFIED	17:23
ALCO TU 29.60 in	.082	17:23
BLANK TEST	.000	17:24
SUBJECT SAMPLE	.000	17:25
BLANK TEST	.000	17:26
ALCO TU 29.60 in	.082	17:26
BLANK TEST	.000	17:27

BMB 10/5/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

SEPTEMBER 30, 2010
TIME 17:19

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.60 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~