VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

Supervisor/Operator Performing the Verification Procedure:

A

Name Christopher M Babcock ID# 3867 Date 3-4-09

Agency King Cove PD Phone # 497-2210

Instrument Location Squad Room

B

Alco S/N X124646 Target Value .079 High Pressure 900psi

Alco Test Values .082 (1st Alco) .082 (2nd Alco)

Signature CDB 3/10/09

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Christopher M Babcock, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

3/11/09

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

MARCH 04, 2009 CWB 3/10/09

OPERATOR'S NAME:
BABCOCK CHRISTOPHER M
OPERATOR'S NUMBER: 3867
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 123456
DEPT/AGENCY: KVC1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X124646

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.20 in

ALCO TARGET	.079	11:07
BLANK TEST	.000	11:08
INTERNAL STANDARD	VERIFIED	11:08
ALCO TV 30.22 in CWB 3/10/09	.082	11:08
BLANK TEST	.000	11:09
SUBJECT SAMPLE CWB	.000	11:10
BLANK TEST	.000	11:10
ALCO TV 30.22 in	.082	11:11
BLANK TEST	.000	11:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

MARCH 04, 2009
TIME 11:04

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.22 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~