

VERIFICATION OF CALIBRATION REPORT
Of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory – statewide Breath Alcohol Program

A **DataMaster cdm S/N 130299**

Supervisor/Operator Performing the Verification Procedure:

Name: Robert Gould ID#: 2489 Date: 3-11-2008 ✓

Agency: King Cove Department of Public Safety Phone #: 497-2210

Instrument Location: Squad Room

B

Alco S/N: X172946 ✓ Target Value: .078 ✓ High Pressure: 450psi

Alco Test Value Average: $\frac{0.081}{.081}$ $\cdot 0.080$ $\frac{0.080}{.080}$ $\cdot 0.081$

1st Alco $\frac{0.081}{.081}$ 3/20/08 2nd Alco $\frac{0.080}{.080}$ 3/20/08

Robert J Gould

(OVER) *CUB 3/19/08*

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert J Gould, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

J Swartz _____ Date 3/25/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

MAR 18 2008

VERIFICATION OF CALIBRATION REPORT
Of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory – statewide Breath Alcohol Program

C

DataMaster cdm S/N 130299

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130299

MARCH 11, 2008 *CSB 3/19/08*

OPERATOR'S NAME: BOULED ROBERT J
 OPERATOR'S NUMBER: 2489
 SUBJECT'S LAST NAME:

VOC
 SUBJECT'S FIRST NAME/MI :
 VOC

O.L. #: VOC
 DEPT/AGENCY: KUC1
 CASE/REPORT: VOC
 TEST TYPE: U
 ALCO TARGET VALUE: .078
 ALCO S/N: X172946 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.30 in
 ALCO TARGET .079 13:56
 BLANK TEST .000 13:57
 INTERNAL STANDARD VERIFIED 13:57
 ALCO TV 30.30 in .001 13:58
 BLANK TEST .000 13:58
 SUBJECT SAMPLE *R56* .000 *CSB 3/19/08* 13:59
 BLANK TEST .000 14:00
 ALCO TV 30.30 in .000 14:00
 BLANK TEST .000 14:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130299

MARCH 11, 2008
 TIME 13:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
 PROGRAM: OKAY
 SOFTWARE DATE: 02/20/01
 HEATERS
 SAMPLE CHAMBER: 49c
 BREATH TUBE: 42c
 BAROMETER: 30.29 in
 FLOW DETECTOR: OKAY
 PUMP
 HIGH SPEED: OKAY
 DETECTOR: OKAY
 FILTERS: OKAY
 QUARTZ STANDARD: OKAY
 CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
 H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
 p q r s t u v w x y z { | } ~ #