VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

Supervisor/Operator Performing the Verification Procedure:

A

Name Christopher M Babcock ID# 3867 Date 4-28-08

Agency King Cove PD Phone # 497-2210

Instrument Location Squad Room

B

Alco S/N ~~X130195~~ X172946 Target Value .078 High Pressure 400

Alco Test Values .078 (1st Alco) .078 (2nd Alco)

Signature CWB 5/5/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that,

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Christopher M Babcock, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

5/5/08

Jeanne Swartz — Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

APRIL 28, 2008 CDB 5/5/08

OPERATOR'S NAME:
BABCOCK CHRISTOPHER M
OPERATOR'S NUMBER: 3867
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: KVC1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: ~~XL30195~~ CDB
X172946

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.40 in		
ALCO TARGET	.079	09:11
BLANK TEST	.000	09:12
INTERNAL STANDARD	VERIFIED	09:13
ALCO TV 30.40 in	.078	09:13
BLANK TEST	.000	09:14
SUBJECT SAMPLE	.000 CDB 5/5/08	09:14
BLANK TEST	.000	09:15
ALCO TV 30.40 in	.078	09:15
BLANK TEST	.000	09:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

APRIL 28, 2008
TIME 09:01

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.38 in
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~