

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130299</u> ✓	
Name	<u>Chris Babcock</u>	ID#	<u>3867</u> Date <u>5-1-09</u> ✓
A Agency	<u>King Cove PD</u>	Phone #	<u>497-2210</u>
Instrument Location	<u>Squad Room</u> ✓		
Alco S/N	<u>X124646</u> <u>130299</u> cas 2/13/09	Target Value	.081 ✓ <u>.079</u> cas
		High Pressure	<u>800</u>
B Alco Test Values	<u>.081</u> ✓ 1 st Alco	<u>.082</u> ✓ 2 nd Alco	
Signature			<u>003</u> <u>5/18/09</u>
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 22 day of May, 2009.


(Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

MAY 01, 2009

OPERATOR'S NAME:
BARCOCK CHRISTOPHER M
OPERATOR'S NUMBER: 3867
SUBJECT'S LAST NAME:

UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 123456
DEPT/AGENCY: KUC1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: X124646

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.96 in
ALCO TARGET .079 14:27
BLANK TEST .000 14:28
INTERNAL STANDARD VERIFIED 14:28
ALCO TU 29.96 in .081 14:29
BLANK TEST .000 14:29
SUBJECT SAMPLE .000 14:30
BLANK TEST .000 14:31
ALCO TU 29.96 in .082 14:31
BLANK TEST .000 14:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

MAY 01, 2009
TIME 14:24

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.96 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL MNOPQRSTU VWXYZ[\]^_`abcde fgh i jk lmnop
qrstuvwxyz{|}~||