

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 09 2009

DataMaster cdm S/N <u>130301</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>Kyle Carrington</u>	ID# <u>5336</u> Date <u>11-3-2009</u> ✓
A Agency <u>UAFPD</u>	Phone # <u>474-7721</u>
Instrument Location <u>612 Yukon Drive Fairbanks, AK 99775</u> ✓	
B Alco S/N <u>X301583</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>800 PSI</u>	
Alco Test Values	<u>.088</u> ✓ 1 st Alco
	<u>.087</u> ✓ 2 nd Alco
Signature	<u>[Handwritten Signature]</u> ✓
(OVER)	

OMB
11/14/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

[Handwritten Signature] (Notary Seal Stamp)
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

NOVEMBER 03, 2009

OPERATOR'S NAME:
CARRINGTON, KYLE
OPERATOR'S NUMBER: 5336
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:

I
O.L. #: 0
DEPT/AGENCY: FBK7
CASE/REPORT: 0
TEST TYPE: U
ALCO TARGET VALUE: .004
ALCO S/N: X301583

BREATH ANALYSIS

.004 ADJUSTED FOR 29.55 in
ALCO TARGET .002 13:58
BLANK TEST .000 13:59
INTERNAL STANDARD VERIFIED 13:59
ALCO TV 29.55 in .008 14:00
BLANK TEST .000 14:01
SUBJECT SAMPLE .000 14:01
BLANK TEST .000 14:02
ALCO TV 29.55 in .007 14:02
BLANK TEST .000 14:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

NOVEMBER 03, 2009
TIME 13:56

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 29.55 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxy{|}~■