VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

Supervisor/Operator Performing the Verification Procedure:

A

Name SGT. KEVIN NELSON ID# 5349 Date 2/4/09

Agency NORTH POLE POLICE DEPARTMENT Phone # (907) 488-0612

Instrument Location NPPD, 125 SNOWMAN LN, NORTH POLE, AK 99705

B

Alco S/N K172988 Target Value .080 High Pressure 800 PSI

Alco Test Values .084 (1st Alco) .083 (2nd Alco)

Signature [signature] 2/18/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

[signature] 2/18/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

FEB 09 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

FEBRUARY 04, 2009 CDB 2/18/09

OPERATOR'S NAME:
NELSON, KEVIN HOON
OPERATOR'S NUMBER: 5349
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 0000000
DEPT/AGENCY: NPLI
CASE/REPORT: 09-000
TEST TYPE: V
ALCO TARGET VALUE: .080
ALCO S/N: X172988

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.44 in

ALCO TARGET	.078	18:40
BLANK TEST	.000	18:41
INTERNAL STANDARD	VERIFIED	18:41
ALCO TV 29.44 in	.084	18:41
BLANK TEST	.000	18:42
SUBJECT SAMPLE CDB 2/18/09	.000	18:43
BLANK TEST	.000	18:44
ALCO TV 29.44 in	.083	18:44
BLANK TEST	.000	18:45

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

FEBRUARY 04, 2009
TIME 18:34

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.44 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~