


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130302</u>
Name	<u>Kevin Nelson</u>	ID#	<u>5349</u> Date <u>4/20/08</u> ✓
A Agency	<u>North Pole Police Department</u> ✓	Phone #	<u>(907)488-0612</u>
Instrument Location <u>NPPD, 125 Snowman Lane, NP, AK 99705</u>			
B Alco S/N	<u>X124634</u> ✓	Target Value	<u>.081</u> High Pressure <u>450 psi</u>
Alco Test Values		<u>.081</u>	<u>.082</u>
		1 st Alco	2 nd Alco
Signature			<u>COB</u> <u>4/28/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/28/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130302
APRIL 20, 2008 *COB 11/23/08*
OPERATOR'S NAME:
NELSON, KEVIN HOON
OPERATOR'S NUMBER: 5349
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
D.L. #: 000000
DEPT/AGENCY: NPL1
CASE/REPORT: 08-000
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X124634

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.99 in		
ALCO TARGET	.081	02:56
BLANK TEST	.000	02:57
INTERNAL STANDARD	VERIFIED	02:57
ALCO TV 29.99 in	.081	02:58
BLANK TEST	.000	02:58
SUBJECT SAMPLE	.000 <i>COB 11/23/08</i>	02:59
BLANK TEST	.000	03:00
ALCO TV 29.99 in	.082	03:00
BLANK TEST	.000	03:01

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATA MASTER cdm 130302
APRIL 20, 2008
TIME 02:54

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJKL
HIJKLMNPQRSTUVWXYZ[\]^_`abcdefgijklmnop
pqrstuvwxyz{|}~