

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 18 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130302</u> ✓
Name	<u>Brandi Barnett</u>	ID# <u>7038</u> Date <u>9/16/09</u> ✓
A Agency	<u>SCDL</u>	Phone# <u>269-5695</u>
Instrument Location	<u>SCDL</u> ✓	
Alco S/N <u>SCOTT08</u> ✓ Target Value <u>0.080</u> ✓ High Pressure <u>600 PSI</u>		
Alco Test Values		
1 <sup>st</sup> Alco	<u>0.077</u> ✓	<u>0.078</u> ✓
2 <sup>nd</sup> Alco		
Signature	<u>Brandi M. Barnett</u> ✓	<u>CMB</u> <u>9/25/09</u>
(OVER)		

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 8<sup>th</sup> day of Oct, 2009.

Carolyn M. Noland  
Carolyn M. Noland

- Notary Public, State of Alaska
- Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

SEPTEMBER 16, 2009 ✓ *COB 9/25/09*

OPERATOR'S NAME:

BARNETT/B

OPERATOR'S NUMBER: 7038

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: AAR1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: 0.090

ALCO S/N: SCOTT08

--- BREATH ANALYSIS ---

.090 ADJUSTED FOR 29.36 in		
ALCO TARGET	.078	10:29
BLANK TEST	.000	10:30
INTERNAL STANDARD	VERIFIED	10:30
ALCO TV 29.36 in	.077	10:30
BLANK TEST	.000	10:31
SUBJECT SAMPLE <i>COB 9/25/09</i>	.000	10:31
BLANK TEST	.000	10:32
ALCO TV 29.36 in	.078	10:33
BLANK TEST	.000	10:33

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

SEPTEMBER 16, 2009 ✓  
TIME 10:23

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	43c
BREATH TUBE:	42c
BAROMETER:	29.38 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~`