

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130302</u> ✓
Name	<u>Kevin Nelson</u>	ID#	<u>5349</u> Date <u>10/21/08</u> ✓
A	Agency	<u>North Pole Police Dept.</u>	Phone # <u>488-6902</u>
	Instrument Location	<u>125 Snowman Lane, North Pole, AK 99705</u> ✓	
B	Alco S/N	<u>1172988</u> ✓	Target Value <u>0.080</u> High Pressure <u>1100psi</u>
	Alco Test Values	<u>0.084</u> ✓ 1 st Alco	<u>0.084</u> ✓ 2 nd Alco
	Signature	<u>[Signature]</u> ✓	

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/2/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

OCTOBER 21, 2008 ✓ 12/21/08

OPERATOR'S NAME:
NELSON, KEVIN HOON
OPERATOR'S NUMBER: 5349
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 000000
DEPT/AGENCY: NPL1
CASE/REPORT: 00-0000
TEST TYPE: V
ALCO TARGET VALUE: .080 ✓
ALCO S/N: X172988

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.59 in		
ALCO TARGET	.079	10:49
BLANK TEST	.000	10:50
INTERNAL STANDARD	VERIFIED	10:50
ALCO TV 29.59 in	.084 ✓	10:50
BLANK TEST	.000	10:51
SUBJECT SAMPLE	.000	10:52
BLANK TEST	.000	10:53
ALCO TV 29.59 in	.084 ✓	10:53
BLANK TEST	.000	10:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130302

OCTOBER 21, 2008 ✓
TIME 10:46

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.59 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; { } < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~