

FEB 02 2010 ²⁰¹¹ _{2/2/11}

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303V

Supervisor/Operator Performing the Verification Procedure:
 Name JAMES I JENSEN, JR ID# 3393V Date 1-29-3011 ²⁰¹¹ _{2/2/11}

A Agency YAKUTAT DEPT. OF PUBLIC SAFETY Phone # 784-3246
 Instrument Location BOILING ROOM AT POLICE DEPT.

B Alco S/N 77286V Target Value .082V High Pressure 250

Alco Test Values .081
.083 ²⁰¹¹
 1st Alco 2/2/2011 2nd Alco .081

Signature [Handwritten Signature]

(OVER)

[Handwritten initials]
3/31/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAR, 2010 ²⁰¹¹ _{2/2/11}

[Handwritten Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130303

BMB
3/31/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

JANUARY 29, 2011

OPERATOR'S NAME:
JENSEN, JR JAMES I
OPERATOR'S NUMBER: 3393
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC
DEPT/AGENCY: YAK1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: 77286

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.43 in
ALCO TARGET .083 19:10
BLANK TEST .000 19:11
INTERNAL STANDARD VERIFIED 19:11
ALCO TV 30.43 in .081 19:11
BLANK TEST .000 19:12
SUBJECT SAMPLE .000 19:13
BLANK TEST .000 19:13
ALCO TV 30.43 in .081 19:14
BLANK TEST .000 19:14

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

JANUARY 29, 2011
TIME 19:09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.45 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~