APR 11 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 13Ø3Ø3

Supervisor/Operator Performing the Verification Procedure:

Name JAMES I JENSEN JR ID# 3393 Date 3-31-11

A

Agency YAKUTAT DPS Phone # 784-32Ø6

Instrument Location AT POLICE STATION (BOOKING ROOM)

B

Alco S/N ZAD Target Value .Ø8Ø High Pressure 95Ø

Alco Test Values .Ø76 (1st Alco) .Ø77 (2nd Alco)

Signature

(OVER)

4/12/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27 day of MAY, 2010. NJB 2011

(Notary Seal Stamp)

~~Daniel L. Nicholas~~ KYLE KLOPFER
Notary Public, State of Alaska
Commission Expires with Office

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

KYLE KLOPFER NOTARY PUBLIC STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303 BMB 4/12/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

MARCH 31, 2011

BMB 4/12/11

OPERATOR'S NAME:
JENSENJAMESI
OPERATOR'S NUMBER: 3393
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: YAK1
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .080
ALCO S/N: 2AD

— BREATH ANALYSIS —

.080 ADJUSTED FOR 29.64 in

ALCO TARGET	.079	01:34
BLANK TEST	.000	01:35
INTERNAL STANDARD	VERIFIED	01:35
ALCO TV 29.64 in	.076	01:35
BLANK TEST	.000	01:36
SUBJECT SAMPLE	.000	01:37
BLANK TEST	.000	01:37
ALCO TV 29.64 in	.077	01:38
BLANK TEST	.000	01:38

BMB 4/12/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

MARCH 31, 2011
TIME 01:32

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	43c
BAROMETER:	29.64 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~