

JUL 23 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

Supervisor/Operator Performing the Verification Procedure:
 Name JAMES I. JENSEN, JR ID# 3393 Date 7-18-11

A Agency YAKUTAT DPS Phone # 784-3206

Instrument Location AT POLICE STATION - BOOKING ROOM
 (Lot 40) BMB

B Alco S/N 07511080A1 Target Value .080 High Pressure 950

Alco Test Values
 1st Alco .079 2nd Alco ~~.080~~ 0.078
BMB
8/1/11

Signature [Handwritten Signature]

(OVER)

BMB
8/1/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug. 2011

[Handwritten Signature] (Notary Seal Stamp)
Daniel L. Nicholas - Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

JULY 18, 2011 *Bmb 8/1/11*

OPERATOR'S NAME: JENSEN, JAMES I
OPERATOR'S NUMBER: 3393
SUBJECT'S LAST NAME: U
SUBJECT'S FIRST NAME/MI: U
O.L. #: U
DEPT/AGENCY: YAK1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .080
ALCO S/N: 0751108 *(4C) Bmb 8/1/11*

— BREATH ANALYSIS —

.080 ADJUSTED FOR 30.03 in		
ALCO TARGET	.080	22:36
BLANK TEST	.000	22:37
INTERNAL STANDARD	VERIFIED	22:37
ALCO TV 30.03 in	.079	22:37
BLANK TEST	.000	22:38
SUBJECT SAMPLE <i>Bmb 8/1/11</i>	.000	22:38
BLANK TEST	.000	22:39
ALCO TV 30.03 in	.078	22:39
BLANK TEST	.000	22:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

JULY 18, 2011
TIME 22:33

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/00/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.03 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■