VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

Supervisor/Operator Performing the Verification Procedure:

A

Name DARYL RECE ID# 2088 Date 02/16/09

Agency SITKA POLECE DEPT. Phone # 747-3245

Instrument Location 304 LAKE STREET

B

Alco S/N 77180 Target Value .078 High Pressure 100

Alco Test Value Average .083 (1st Alco) .083 (2nd Alco)

Signature Daryl Rice

8/23/09

Under the Alaska Rules of Evidence, I certify that; (OVER)

(Do Not Write in this area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daryl Rice, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

02/23/09

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

FEBRUARY 16, 2009 HE PST 100 Daryl Rice 2/16/09

OPERATOR'S NAME:
RICE DARYL DEAN
OPERATOR'S NUMBER: 2088
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI : 2/23/09
VOC
O.L. #: VOC
DEPT/AGENCY: SIT1
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: 77180

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.99 in

ALCO TARGET	.078	22:22
BLANK TEST	.000	22:23
INTERNAL STANDARD	VERIFIED	22:23
ALCO TV 29.99 in	.083	22:23
BLANK TEST	.000	22:24
SUBJECT SAMPLE 2/23/09	.000	22:25
BLANK TEST	.000	22:26
ALCO TV 29.99 in	.083	22:26
BLANK TEST	.000	22:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

FEBRUARY 16, 2009 HE PST 100 Daryl Rice 2/16/09
TIME 22:20

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	43c
BAROMETER:	29.99 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~