

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 14 2009

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130303 ✓

Name Ben Dudley ID# 3532 Date 8.7.09 ✓

A Agency Wasilla PD Phone # 352.5408

Instrument Location 1800 E. Parks Hy. Wasilla, AK (Wasilla PD)

B Alco S/N 70108 Target Value .087 ✓ High Pressure 1200

Alco Test Values .086 ✓ .085 ✓
1st Alco 2nd Alco

Signature Ben Dudley CBS
8/27/09


(OVER)

- (Do Not write in the area below)
- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

AUGUST 07, 2009

OPERATOR'S NAME:

DUDLEY/BEN/H

OPERATOR'S NUMBER: 3532

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

L

O.L. #: L

DEPT/AGENCY: WAS1

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .087 ✓

ALCO S/N: 70108

— BREATH ANALYSIS —

| | | |
|----------------------------|----------|-------|
| .087 ADJUSTED FOR 29.88 in | | |
| ALCO TARGET | .086 | 07:02 |
| BLANK TEST | .000 | 07:03 |
| INTERNAL STANDARD | VERIFIED | 07:03 |
| ALCO TV 29.78 in | .086 ✓ | 07:03 |
| BLANK TEST | .000 | 07:04 |
| SUBJECT SAMPLE | .000 | 07:04 |
| BLANK TEST | .000 | 07:05 |
| ALCO TV 29.80 in | .085 ✓ | 07:05 |
| BLANK TEST | .000 | 07:06 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303

AUGUST 07, 2009 ✓

TIME 06:59

— DIAGNOSTIC CHECK —

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 01/08/09 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.88 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~■