

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 09 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130303

Name JEAN ACHEE ID# 1470 Date 10-1-2009

A Agency WASILLA POLICE DEPT. Phone # 352-5401

Instrument Location WASILLA POLICE DEPT.

B Alco S/N 70108 Target Value .087 High Pressure 600

Alco Test Value Average .086 .086  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Jean Achee

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

SMB  
11/14/09

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 3<sup>rd</sup> day of Dec, 2009.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

~~130570~~ DW 11-10-09  
130303

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303 ✓

OCTOBER 01, 2009 ✓

OPERATOR'S NAME: *J.P.*

ACHEE, JEAN P

OPERATOR'S NUMBER: 1470

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567890

DEPT/AGENCY: WAS1

CASE/REPORT: 09-1001

TEST TYPE: V

ALCO TARGET VALUE: .087 ✓

ALCO S/N: 70108 ✓

### --- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.90 in

ALCO TARGET .086

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.88 in .086 ✓

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 29.90 in .086 ✓

BLANK TEST .000

*BMB*  
11/14/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303 ✓

OCTOBER 01, 2009 ✓  
TIME 17:38

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

17: DETECTOR: OKAY

17: FILTERS: OKAY

17: QUARTZ STANDARD: OKAY

17: CALIBRATION: OKAY

17: :

PRINTER TEST

! " \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @  
[ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ Δ