

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130305</u>
Name		<u>Daryl Rice</u>	ID# <u>2088</u> Date <u>10/18/08</u>
A	Agency	<u>Sitka Police Dept</u>	Phone # <u>747-3245</u>
Instrument Location		<u>304 Lake Street</u>	
Alco S/N		<u>77180</u>	Target Value <u>0.078</u> High Pressure <u>600</u>
B	Alco Test Values	<u>0.083</u> 1 <sup>st</sup> Alco	<u>0.083</u> 2 <sup>nd</sup> Alco
Signature		<u>Daryl Rice</u>	<u>CRS</u> <u>11/4/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daryl Rice, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

11/5/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

OCTOBER 18, 2008

OPERATOR'S NAME:  
RICE DARVL DEAN  
OPERATOR'S NUMBER: 2038  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: VOC  
DEPT/AGENCY: SIT1  
CASE/REPORT: VOC  
TEST TYPE: V  
ALCO TARGET VALUE: .078 ✓  
ALCO S/N: 77180 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.12 in  
ALCO TARGET .078 10:26  
BLANK TEST .000 10:27  
INTERNAL STANDARD VERIFIED 10:27  
ALCO TV 30.11 in .083 ✓ 10:28  
BLANK TEST .000 10:29  
SUBJECT SAMPLE .000 10:29  
BLANK TEST .000 10:30  
ALCO TV 30.12 in .083 ✓ 10:31  
BLANK TEST .000 10:31

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130303

OCTOBER 18, 2008  
TIME 10:32

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.12 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghi.jklmno  
pqrstuvwxyz{|}~■