

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 6 3 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130303 ✓

Name Ben H. Dudley ID# 3532 Date 11.24.09 ✓

A Agency Wasilla PD Phone # 907.352.5408

Instrument Location Wasilla PD (1800 E Parks Hwy) Wasilla, AK ✓

B Alco S/N 75509 Target Value .079 High Pressure 1200

Alco Test Value Average .076 .080 ✓
1st Alco .076 2nd Alco .080

Signature Ben H. Dudley

(OVER)

BMB
12/31/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Dec., 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130303 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303 ✓

NOVEMBER 24, 2009 ✓

OPERATOR'S NAME:

DUDLEY/BEN/H

OPERATOR'S NUMBER: 3532

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: WAS1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 75509 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.15 in ✓

ALCO TARGET	.076	00:09
BLANK TEST	.000	00:10
INTERNAL STANDARD	VERIFIED	00:10
ALCO TV 29.15 in	.080 ✓	00:11
BLANK TEST	.000	00:11
SUBJECT SAMPLE	.000	00:12
BLANK TEST	.000	00:13
ALCO TV 29.16 in	.080 ✓	00:13
BLANK TEST	.000	00:14

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130303 ✓

NOVEMBER 24, 2009 ✓

TIME 00:07

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.15 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~Δ