

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130305

Supervisor / Operator Performing the Verification Procedure:

Name: William A. Mills ID#: 2688 Date: February 22, 2012

A. Agency: Hoonah Police Department Phone #: 945-3655

Instrument Location: 379 Raven Drive, Hoonah

Alco S/N: X301532 Target Value: .082 High Pressure: 100 psi

B. Alco Test Value Average: .086 .087
 1st Alco 2nd Alco

Signature William A. Mills

(OVER)

SMB
4/6/12

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 6 day of April, 2012. *4/6/12*

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

100305
BMB 130305
4/6/12

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

FEBRUARY 22, 2012

OPERATOR'S NAME: HILLS WILLIAM
OPERATOR'S NUMBER: 2688
SUBJECT'S LAST NAME:
VERIFICATION OF
SUBJECT'S FIRST NAME/MI :
CALIBRATION
D.L. #: 1234567890
DEPT/AGENCY: HNH
CASE/REPORT: 1234567890
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: X301532

— BREATH ANALYSIS —

.082 ADJUSTED FOR 29.68 in
ALCO TARGET .081 01:17
BLANK TEST .000 01:18
INTERNAL STANDARD VERIFIED 01:18
ALCO TO 29.68 in .086 01:19
BLANK TEST .000 01:19
SUBJECT SAMPLE .000 01:21
BLANK TEST .000 01:21
ALCO TO 29.68 in .087 01:22
BLANK TEST .000 01:22

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

FEBRUARY 22, 2012
TIME 01:24

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 40c
BAROMETER: 29.68 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxy{z{|}~