

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

NOV 27 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130305</u>
Name <u>Anthony Wallace</u>		ID# <u>6555</u> Date <u>11/23/09</u>
A	Agency <u>Hoonah Police Dept</u>	Phone # <u>945-3655</u>
Instrument Location <u>Hoonah Police Dept</u>		
Alco S/N <u>X30153Z</u>		Target Value <u>.082</u> High Pressure <u>800 PSI</u>
B	Alco Test Value Average <u>.086</u>	<u>.086</u>
	1 st Alco	2 nd Alco
Signature <u>[Signature]</u>		
(OVER)		

gmb
1/5/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows (Do Not write in this area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of January, 2010
[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

NOVEMBER 23, 2009 ✓ *BMB 115/10*

OPERATOR'S NAME:
WALLACE ANTHONY M
OPERATOR'S NUMBER: 6555
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/NI :
UOC

O.L. #: UOC
DEPT/AGENCY: HNH
CASE/REPORT: UOC
TEST TYPE: U ✓
ALCO TARGET VALUE: .082 ✓
ALCO S/N: X301532 ✓

— BREATH ANALYSIS —

.082 ADJUSTED FOR 30.01 in		
ALCO TARGET	.082	22:32
BLANK TEST	.000	22:33
INTERNAL STANDARD	VERIFIED	22:33
ALCO TV 30.01 in	.086 ✓	22:33
BLANK TEST	.000	22:34
SUBJECT SAMPLE <i>(circled)</i>	.000	22:35
BLANK TEST	.000 ✓	22:35
ALCO TV 30.01 in	.086	22:36
BLANK TEST <i>(circled)</i>	.000	22:36

BMB 115/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

NOVEMBER 23, 2009 ✓
TIME 22:25

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 40c
BREATH TUBE: 41c
BAROMETER: 30.01 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMN0PQRSTUWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~||