

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | |
|---|--|
| DataMaster cdm S/N <u>130305</u> ✓ | |
| Supervisor/Operator Performing the Verification Procedure: | |
| Name <u>ANDRE ACHEE</u> | ID# <u>2385</u> Date <u>11/25/2008</u> ✓ |
| A | |
| Agency <u>BETHEL POLICE DEPARTMENT</u> | Phone # <u>907-543-7781</u> |
| Instrument Location <u>SQUAD ROOM</u> ✓ | |
| B | |
| Alco S/N <u>X172939</u> ✓ | Target Value <u>.081</u> ✓ High Pressure <u>10</u> |
| Alco Test Values | <u>.081</u> ✓ <u>.080</u> ✓ |
| | 1 st Alco 2 nd Alco |
| Signature  | <u>COB</u> <u>12/23/08</u> |
| (OVER) | |

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Andre Achee, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed





Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

12/23/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

NOVEMBER 25, 2008

OPERATOR'S NAME: *CAB 12/23/08*
ACHEE/ANDRE/A
OPERATOR'S NUMBER: 2385
SUBJECT'S LAST NAME:

TEST
SUBJECT'S FIRST NAME/MI :
VOR

O.L. #: XXXXXX
DEPT/AGENCY: BET1
CASE/REPORT: XXXXXX
TEST TYPE: T
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X172939

--- BREATH ANALYSIS ---

| | | |
|------------------------------------|----------|-------|
| .081 ADJUSTED FOR 29.65 in | | |
| ALCO TARGET | .080 | 15:38 |
| BLANK TEST | .000 | 15:39 |
| INTERNAL STANDARD | VERIFIED | 15:39 |
| ALCO TV 29.65 in | .081 ✓ | 15:40 |
| BLANK TEST | .000 | 15:41 |
| SUBJECT SAMPLE <i>CAB 12/23/08</i> | .000 | 15:41 |
| BLANK TEST | .000 | 15:42 |
| ALCO TV 29.65 in | .080 ✓ | 15:42 |
| BLANK TEST | .000 | 15:43 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

NOVEMBER 25, 2008 ✓
TIME 15:44

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 40c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.65 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrsstuvwxyz{|}~■