

VERIFICATION OF CALIBRATION REPORT

FEB 12 2010

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130306</u> ✓
Name <u>C. Rivala</u>	ID# <u>3791</u>	Date <u>2/11/10</u> ✓
Agency <u>APD</u>	Phone # <u>786-2640</u>	
Instrument Location <u>Floor 203</u> ✓		
Alco S/N <u>X124650</u> ✓		Target Value <u>079</u> ✓
		High Pressure <u>300</u>
Alco Test Values		
<u>075</u> ✓ 1 st Alco	<u>076</u> ✓ 2 nd Alco	
Signature <u>[Signature]</u> / 1349		

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of February, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130306 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306 ✓

FEBRUARY 11, 2010

OPERATOR'S NAME:
RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

A

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124650

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.31 in ✓
ALCO TARGET .077 23:50
BLANK TEST .000 23:51
INTERNAL STANDARD VERIFIED 23:51
ALCO TU 29.33 in .075 23:51
BLANK TEST .000 23:52
SUBJECT SAMPLE .000 23:53
BLANK TEST .000 23:54
ALCO TU 29.33 in .076 23:54
BLANK TEST .000 23:55

2/13/10
2/11/10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306 ✓

FEBRUARY 11, 2010 ✓
TIME 23:55

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.33 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKL*NOPQRSTUUVWXYZ[\]^_`abcde fghi jk lmnop
qrs tuvwxyz{|}~

2/13/10
2/11/10