VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 1 8 2009

DataMaster cdm S/N 130306

Supervisor/Operator Performing the Verification Procedure:

A

Name C. RITALA ID# 3791 Date 2/16/09

Agency A.P.D. Phone # 786-2640

Instrument Location FLORA SUB STATION (4TH AVE)

B

Alco S/N X172933 Target Value .069 High Pressure 450

Alco Test Value Average .070 (1st Alco) .071 (2nd Alco)

Signature /1349 08/23/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

2/23/09

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

FEBRUARY 16, 2009

CAB 2/23/09

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
NOW AT THE FLORA SUB
O.L. #: A
DEPT/AGENCY: ANC1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .069
ALCO S/N: X172933

--- BREATH ANALYSIS ---

.069 ADJUSTED FOR 29.65 in

ALCO TARGET	.068	18:13
BLANK TEST	.000	18:15
INTERNAL STANDARD	VERIFIED	18:15
ALCO TV 29.65 in	.070	18:15
BLANK TEST	.000	18:16
SUBJECT SAMPLE	.000	18:16
BLANK TEST	.000	18:17
ALCO TV 29.67 in	.071	18:18
BLANK TEST	.000	18:18

CAB 2/23/09

CR/1349
2/16/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

FEBRUARY 16, 2009
TIME 18:20

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.67 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

CR/1349
2/16/09