

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

FEB 18 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130306 ✓
A	Name	C. RITALA	ID# 3791 Date 2/16/09 ✓
	Agency	A.P.D.	Phone # 786-2640
	Instrument Location	Mt. View SUB STATION (MIZELLE SUB)	✓
B	Alco S/N	X172933 ✓	Target Value .069 ✓ High Pressure 450
	Alco Test Value Average	.071 ✓	.071 ✓
		1 st Alco	2 nd Alco
	Signature	 /1349 ✓	 2/23/09

(OVER)


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

02/23/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

FEBRUARY 16, 2009 ✓

OPERATOR'S NAME:
RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

MOVING TO 4TH AVE SUB

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: .069 ✓

ALCO S/N: X172933 ✓

--- BREATH ANALYSIS ---

.069 ADJUSTED FOR 29.60 in
ALCO TARGET .068 17:40
BLANK TEST .000 17:41
INTERNAL STANDARD VERIFIED 17:41
ALCO TV 29.60 in .071 ✓ 17:41
BLANK TEST .000 17:42
SUBJECT SAMPLE .000 17:43
BLANK TEST .000 17:44
ALCO TV 29.60 in .071 ✓ 17:44
BLANK TEST .000 17:45

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

FEBRUARY 16, 2009 ✓
TIME 17:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.60 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLNNOPQRSTUWVWXYZ[\]^_`abcdefg h i j k l m n o
p q r s t u v w x y z { | } ~

CR/1349 2/16/09