

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 14 2008

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130306</u> ✓	
Name	<u>D.R. STURKIE</u>	ID#	<u>3559</u> Date <u>05-14-08</u> ✓
Agency	<u>APD</u>	Phone #	<u>786-8900</u>
Instrument Location	<u>Seeley SUBSTATION (SPENARD) ✓</u>		
Alco S/N	<u>75602</u> ✓	Target Value	<u>0.074</u> ✓
		High Pressure	<u>1000</u> (1000)
Alco Test Value Average	<u>.070</u> ✓ 1 st Alco		<u>.074</u> ✓ 2 nd Alco
Signature	<u>D.R. Sturkie</u>		<u>CRS</u> <u>6/6/08</u>

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Daniel Sturkie, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/6/08

Date

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MAY 14, 2008

OPERATOR'S NAME:

STURKIE

OPERATOR'S NUMBER: 3559

SUBJECT'S LAST NAME:

L

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: ANCI

CASE/REPORT: L

TEST TYPE: V

ALCO TARGET VALUE: .074

ALCO S/N: 75602

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.49 in		
ALCO TARGET	.072	08:38
BLANK TEST	.000	08:39
INTERNAL STANDARD	VERIFIED	08:39
ALCO TV 29.49 in	.070	08:39
BLANK TEST	.000	08:40
SUBJECT SAMPLE	.000	08:41
BLANK TEST	.000	08:42
ALCO TV 29.49 in	.074	08:43
BLANK TEST	.000	08:44

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MAY 14, 2008

TIME 08:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.49 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJKLMO
PQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~