

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130306</u> ✓	
Name <u>Patrick E. Martin</u>	ID# <u>4072</u> Date <u>5-16-09</u> <u>7-14-09</u> <i>CMJ 7-14-09</i>
A Agency <u>APD</u>	Phone # <u>786-2601</u>
Instrument Location <u>APD 7-14-09 CMJ</u> ✓	
B Alco S/N <u>X172946</u> ✓ Target Value <u>.082</u> ✓ High Pressure _____	
Alco Test Value Average <u>.079</u> ✓ <u>.080</u> ✓	
1 st Alco 2 nd Alco	
Signature <u>PEM</u> ✓	
COB 7/13/09	


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(Do Not Write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of July, 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MAR 16, 2009 ✓

OPERATOR'S NAME: *CPB 7/13/09*
MARTIN, PATRICK E
OPERATOR'S NUMBER: 4072
SUBJECT'S LAST NAME:
UC
SUBJECT'S FIRST NAME/NI:
U

O.L. #: 1234567
DEPT/AGENCY: ANCI
CRIME/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: X172946

— BREATH ANALYSIS —

.082 ADJUSTED FOR 30.17 in		
ALCO TARGET	.082	04:05
BLANK TEST	.000	04:04
INTERNAL STANDARD	VERIFIED	04:04
ALCO TV 30.17 in	.079 ✓	04:04
BLANK TEST	.000	04:05
SUBJECT SAMPLE	.000 ✓	04:06
BLANK TEST	.000 ✓	04:07
ALCO TV 30.17 in	.000	04:07
BLANK TEST	.000	04:08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

MAR 16, 2009 ✓
TIME: 03:58

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
RAM: OKAY
BATTERY DATE: 02/20/01
METERS
LE CHAMBER: 49c
TUBE: 42c
METER: 30.17 in
DETECTOR: OKAY
SPEED: OKAY
DETECTOR: OKAY
METERS: OKAY
STANDARD: OKAY
METER: OKAY

PRINTER TEST

123456789:;<?@BCDEFG
"NOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~