

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130306</u>
Name		<u>R. Blanton</u>	ID# <u>25084</u> Date <u>7-4-08</u> ✓
A	Agency	<u>APD</u>	Phone # _____
Instrument Location <u>Seely Substation</u> ✓			
Alco S/N		<u>57447</u> ✓	Target Value <u>.071</u> ✓ High Pressure <u>1200</u>
B	Alco Test Values		
		<u>.076</u> ✓ 1 <sup>st</sup> Alco	<u>.075</u> ✓ 2 <sup>nd</sup> Alco
Signature		<u>R. Blanton</u>	<u>CS</u> <u>7/10/08</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that, (Print name) \_\_\_\_\_ (Type name in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert Blanton, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

7/13/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

JULY 04, 2008

OPERATOR'S NAME:

BLANTON, R

OPERATOR'S NUMBER: 5004

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

A

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .071 ✓

ALCO S/N: 57447 ✓

### --- BREATH ANALYSIS ---

.071 ADJUSTED FOR 29.72 in  
ALCO TARGET .070 16:36  
BLANK TEST .000 16:37  
INTERNAL STANDARD VERIFIED 16:37  
ALCO TV 29.72 in .076 ✓ 16:38  
BLANK TEST .000 16:39  
SUBJECT SAMPLE .000 16:39  
BLANK TEST .000 16:40  
ALCO TV 29.72 in .075 16:40  
BLANK TEST .000 16:41

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

JULY 04, 2008 ✓

TIME 16:44

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 41c  
BAROMETER: 29.72 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! "\$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢