

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130306 ✓			
Name	C. RITALA	ID#	3791	Date	7/21/08 ✓	
A	Agency	ANCHORAGE P.D.	Phone #			
	Instrument Location	Seely Substation (SPENARD) ✓				
B	Alco S/N	79966 ✓	Target Value	079 ✓	High Pressure	1100
	Alco Test Values	.078 ✓ 1 st Alco	.078 ✓ 2 nd Alco			
	Signature	Chris Ritala / 1349				CRB 7/29/08
(OVER)						


Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

7/30/08
Date

JUL 22 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

JULY 21, 2008

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/NI :

A

O.L. #: A

DEPT/AGENCY: ANCI

CASE/REPORT: A

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 78966 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.91 in
ALCO TARGET .078 21:50
BLANK TEST .000 21:51
INTERNAL STANDARD VERIFIED 21:51
ALCO TV 29.91 in .078 ✓ 21:51
BLANK TEST .000 21:52
SUBJECT SAMPLE .000 21:53
BLANK TEST .000 21:53
ALCO TV 29.91 in .078 ✓ 21:54
BLANK TEST .000 21:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

JULY 21, 2008 ✓

TIME 21:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.91 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKL;NOPQRSTUVWXYZ\]^_`abcdefg|hijklmno
pqrstuvwxyz{|}~