


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130306 ✓
Name	Swaran Singh	ID#	4945 26734 ✓
Agency	APD	Date	12/20/08 ✓
Instrument Location	Sully SOB	Phone #	12-2208 CYP 786-8256
Alco S/N	X 172966 ✓	Target Value	.078 ✓
		High Pressure	1150 PSI
Alco Test Value Average	.078 ✓		.078 ✓
	1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature			COS 12/23/08


(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Swaran Singh, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

12/23/08  
Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

*JA*  
*26734*  
*12/20/08*

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

DECEMBER 20, 2008  
*CAB*  
*12/23/08*

OPERATOR'S NAME:  
SINGH/SUNAR/ANSON  
OPERATOR'S NUMBER: 4945  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL.  
SUBJECT'S FIRST NAME/MI :  
L  
O.L. #: L  
DEPT/AGENCY: ANCI  
CASE/REPORT :  
TEST TYPE: V  
ALCO TARGET VALUE: .075 ✓  
ALCO S/N: X172966 ✓

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 30.25 in		
ALCO TARGET	.075 ✓	23:28
BLANK TEST	.000	23:29
INTERNAL STANDARD	VERIFIED	23:29
ALCO TO 30.25 in	.075 ✓	23:29
BLANK TEST	.000	23:30
SUBJECT SAMPLE	.000 ✓	23:31
BLANK TEST	.000	23:32
ALCO TO 30.25 in	.075 ✓	23:32
BLANK TEST	.000	23:33

*CAB*  
*12/23/08*

## DIAGNOSTIC CHECK

*JA*  
*26734*  
*12/20/08*

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

DECEMBER 20, 2008 ✓  
TIME 23:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.25 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
"##%&'()\*+,-./:0123456789:;=<?@ABCDEFB  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstvwxyz{|}~&#