

JAN 01 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130306

Name Derek Sitz ID# 4096 Date 12.29.07

A Agency Anchorage Police Department Phone # 764.9607

Instrument Location Seely Substation

B Alco S/N X172966 Target Value .084 High Pressure 1200

Alco Test Values	<u>.0808</u>	<u>.079</u>
	1 st Alco	2 nd Alco

Signature Derek Sitz

(OVER)

Under the Alaska Rules of Evidence, I certify that (in the area below)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derek Sitz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 1/17/08
 Jeanne Swartz Date
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

DECEMBER 29, 2007

OPERATOR'S NAME:
SITZ, DEREK W
OPERATOR'S NUMBER: 4096
SUBJECT'S LAST NAME:

UOC
SUBJECT'S FIRST NAME/MI:

UOC
D.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: 00-1234
TEST TYPE: U
ALCO TARGET VALUE: .084
ALCO S/N: X172966

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.52 in

ALCO TARGET	.082	00:32
BLANK TEST	.000	00:33
INTERNAL STANDARD	VERIFIED	00:33
ALCO TV 29.52 in	.080	00:34
BLANK TEST	.000	00:35
SUBJECT SAMPLE	.000	00:36
BLANK TEST	.000	00:36
ALCO TV 29.52 in	.079	00:37
BLANK TEST	.000	00:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

DECEMBER 29, 2007
TIME 00:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.52 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

#####