VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306 ✓

Supervisor/Operator Performing the Verification Procedure:

A

Name C Ritala ID# 3791 Date 12/29/08 ✓

Agency APD Phone # 786-2640

Instrument Location Seely Sub Station ✓

B

Alco S/N X172966 ✓ Target Value .078 ✓ High Pressure 900

Alco Test Values .078 (1st Alco) .07 (2nd Alco)

Signature [signature] 1349 ✓ 12/30/08

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 12/31/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

DEC 29 2008

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130306 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

DECEMBER 29, 2008 — CDB 12/30/08

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: ANC1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .078 ✓
ALCO S/N: X172966

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.96 in

ALCO TARGET	.078	17:11
BLANK TEST	.000	17:12
INTERNAL STANDARD	VERIFIED	17:12
ALCO TV 29.96 in	.078 ✓	17:13
BLANK TEST	.000	17:13
SUBJECT SAMPLE CDB 12/30/08	.000	17:14
BLANK TEST	.000	17:15
ALCO TV 29.98 in	.077 ✓	17:16
BLANK TEST	.000	17:17

CR/1349 12/29/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

DECEMBER 29, 2008 ✓
TIME 17:05

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	29.96 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgnijklmno
pqrstuvwxyz{|}~

CR/1349 12/29/08