

FEB 15 2011
2011
3/16/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130307 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name Fred J. Whittingham ID# 7286 Date 2/11/11 ✓
 A Agency Whittier PD Phone # 242-9553
 Instrument Location Whittier PD

B Alco S/N X124654 ✓ Target Value 0.001 ✓ High Pressure 650
 Alco Test Values
 1st Alco 0.077 BMB 3/6/11
 2nd Alco 0.077 BMB 3/6/11
 Signature [Signature]

(OVER)

(Do Not write in the area below)

BMB
3/6/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of MAR, 2011

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130307

BMB
3/16/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

FEBRUARY 11, 2011

OPERATOR'S NAME: *BMB*
WHITTINGHAM, FRED
OPERATOR'S NUMBER: 7286
SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
UO

O.L. #: UOC

DEPT/AGENCY: WTI1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .001

ALCO S/N: X124656

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.60 in
ALCO TARGET .000 10:12
BLANK TEST .000 10:13
INTERNAL STANDARD VERIFIED 10:13
ALCO TU 29.60 in .077 10:14
BLANK TEST .000 10:15
SUBJECT SAMPLE *BMB* .000 10:15
BLANK TEST *3/16/11* .000 10:16
ALCO TU 29.60 in .077 10:16
BLANK TEST .000 10:17

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

FEBRUARY 11, 2011
TIME 10:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 51c
BREATH TUBE: 41c
BAROMETER: 29.60 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~!@