

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>Fred Whittingham</u>	ID# <u>7286</u> Date <u>4/11/11</u>
Agency <u>Whittier Police Dept</u>	Phone # _____
Instrument Location <u>Whittier PD</u>	
Alco S/N <u>X124656</u> Target Value <u>.081</u> High Pressure <u>650</u>	
Alco Test Values	
<u>0.080 BMB</u> <u>.007</u> 1st Alco	<u>.079</u> 2 <sup>nd</sup> Alco
Signature <u>Fred Whittingham</u>	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2011.

Kyle Klopfer  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



BMB  
4/13/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130307 *BMB*  
4/13/11

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130307

APRIL 11, 2011

OPERATOR'S NAME: *BMB*  
WHITTINGHAM, FRED J  
OPERATOR'S NUMBER: 7286  
SUBJECT'S LAST NAME: *4/13/11*

UC  
SUBJECT'S FIRST NAME/MI :  
UC

D.L. #: UC  
DEPT/AGENCY: WITI  
CASE/REPORT: UC  
TEST TYPE: U  
ALCO TARGET VALUE: .001  
ALCO S/N: X124656

### --- BREATH ANALYSIS ---

.001 ADJUSTED FOR 30.03 in  
ALCO TARGET .001 14:08  
BLANK TEST .000 14:08  
INTERNAL STANDARD VERIFIED 14:09  
ALCO TV 30.03 in .000 14:09  
BLANK TEST .000 14:10  
SUBJECT SAMPLE *BMB* .000 14:10  
BLANK TEST *4/13/11* .000 14:11  
ALCO TV 30.03 in .079 14:11  
BLANK TEST .000 14:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130307

APRIL 11, 2011  
TIME 14:03

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 50c  
BREATH TUBE: 41c  
BAROMETER: 30.03 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJ KLMNOPQRSTU VWXYZ[\]^\_`abc defghijklmno  
pqrstuvwxy z{|}~@