

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

OCT 26 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130307 ✓
Name JAMES I. JENSEN, JR ID# 3393 Date 10-13-09 ✓
A Agency YAKUTAT DPS Phone # 784-3206
Instrument Location AT NEW POLICE STATION (FROM OLD STATION) ✓
B Alco S/N 77286 ✓ Target Value 0.082 High Pressure 700
Alco Test Values 0.082 ✓ 0.083 ✓
1st Alco 11/9/09 ✓ 2nd Alco 11/9/09 ✓
Signature [Signature]
(OVER)

(Do Not write in the area below)

BMB
11/14/09

Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 3rd day of Dec, 2009.

[Signature] (Notary Seal Stamp)

Carolynn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 13φ3φ7 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307 ✓

OCTOBER 13, 2009 ✓

OPERATOR'S NAME: JENSEN, JAMES I
OPERATOR'S NUMBER: 3393
SUBJECT'S LAST NAME: U
SUBJECT'S FIRST NAME/MI: U
O.L. #: U
DEPT/AGENCY: WAK1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .002 ✓
ALCO S/N: 77286

BMB 11/14/09

— BREATH ANALYSIS —

.002 ADJUSTED FOR 29.83 in		
ALCO TARGET	.001	23:01
BLANK TEST	.000	23:02
INTERNAL STANDARD	VERIFIED	23:02
ALCO TV 29.83 in	.003	23:02
BLANK TEST	.000	23:03
SUBJECT SAMPLE	.000	23:04
BLANK TEST	.000	23:05
ALCO TV 29.83 in	.002	23:05
BLANK TEST	.000	23:06

BMB 11/14/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307 ✓

OCTOBER 13, 2009 ✓
TIME 22:55

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 42c

BAROMETER: 29.83 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢