

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130307</u>	
Name <u>JAMES I. JENSEN, JR</u>		ID# <u>3393</u>	Date <u>11-4-08</u>
A	Agency <u>YAKUTAT DPS</u>	Phone # <u>784-3206</u>	
Instrument Location <u>AT POLICE STATION</u>			
Alco S/N <u>X124635</u>		Target Value <u>.079</u>	High Pressure <u>1005</u>
B	Alco Test Values	<u>.079</u> ✓	<u>.079</u> ✓
		1 st Alco	2 nd Alco
Signature <u>[Handwritten Signature]</u>		<u>CS</u> <u>11/18/08</u>	

(OVER)

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James I Jensen, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Handwritten Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/19/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130307

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

NOVEMBER 04, 2008

OPERATOR'S NAME:

JENSEN, JAMES I

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: YAKI

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124635

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.07 in
ALCO TARGET .079 19:27
BLANK TEST .000 19:28
INTERNAL STANDARD VERIFIED 19:28
ALCO TV 30.07 in .079 ✓ 19:28
BLANK TEST .000 19:29
SUBJECT SAMPLE .000 19:30
BLANK TEST .000 19:31
ALCO TV 30.06 in .079 ✓ 19:31
BLANK TEST .000 19:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

NOVEMBER 04, 2008

TIME 19:21

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 30.07 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢