

DEC 29 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130 307

Supervisor/Operator Performing the Verification Procedure:

Name Fred Whittingham ID# 7286 Date 12/23/10

A Agency Whittier Police Phone # 472-2340

Instrument Location Whittier PD

B Alco S/N X124656 Target Value .081 High Pressure 750

Alco Test Values .081 .080
 1st Alco 2nd Alco

Signature [Handwritten Signature] CSB
12/30/10

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 9th day of MAR, 2010

[Handwritten Signature]
 (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130307
COB 12/29/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

DECEMBER 23, 2010 *COB 12/30/10*

OPERATOR'S NAME:
WHITTINGHAM, FRED
OPERATOR'S NUMBER: 7286
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: MPD
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .081
ALCO S/N: X124656

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.19 in		
ALCO TARGET	.081	10:13
BLANK TEST	.000	10:14
INTERNAL STANDARD	VERIFIED	10:14
ALCO TV 30.19 in	.081	10:15
BLANK TEST	.000	10:16
SUBJECT SAMPLE	.000	10:16
BLANK TEST	.000	10:17
ALCO TV 30.19 in	.000	10:17
BLANK TEST	.000	10:18

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

DECEMBER 23, 2010
TIME 10:10

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.19 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghi.jklmno
pqrstuvwxyz{|}~`