VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130308

Supervisor/Operator Performing the Verification Procedure:

Name Malik Jones ID# 5830 Date 5/14/08

A

Agency AST Phone # 832-5554

Instrument Location Palmer

Alco S/N X173355 Target Value .082 High Pressure 850 PSI

B

Alco Test Value Average .075 .076

1st Alco 2nd Alco

Signature

6/3/08

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator Malik Jones whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

6/5/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130308

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130308

MAY 14, 2008 03/03/08

OPERATOR'S NAME:
JONES/MALIK/JAMAL
OPERATOR'S NUMBER: 5830
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
0
O.L. #: 0
DEPT/AGENCY: HLV0
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X173355

— BREATH ANALYSIS —

.082 ADJUSTED FOR 28.06 in

ALCO TARGET	.076	12:27
BLANK TEST	.000	12:29
INTERNAL STANDARD	VERIFIED	12:29
ALCO TV 28.06 in	.075	12:29
BLANK TEST	.000	12:30
SUBJECT SAMPLE	.000	12:31
BLANK TEST	.000	12:31
ALCO TV 28.06 in	.076	12:32
BLANK TEST	.000	12:33

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130308

MAY 14, 2008
TIME 12:36

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	28.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~