

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309 ✓

Supervisor/Operator Performing the Verification Procedure:

Name P. Nelson ID# 4389 ✓ Date 1/9/11 ✓

A Agency AST Phone # 746-9108

Instrument Location MATSU WEST AST

B Alco S/N X172938 ✓ Target Value .076 ✓ High Pressure 1000 PSI

Alco Test Values .076 ✓ .076 ✓
 1st Alco 2nd Alco

Signature [Signature]

(OVER)

[Handwritten initials]
2/8/11


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 2ND day of MAR, 2011

[Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

JANUARY 09, 2011

OPERATOR'S NAME: *BMB*
NELSON PATRICK S
OPERATOR'S NUMBER: 4389
SUBJECT'S LAST NAME: *2/8/11*
VERIFICATION OF
SUBJECT'S FIRST NAME/MI :
CALIBRATION
O.L. #: 12-34567
DEPT/AGENCY: PLHD
CASE/REPORT: 12-34567
TEST TYPE: V
ALCO TARGET VALUE: .076
ALCO S/N: X172938

--- BREATH ANALYSIS ---

.076 ADJUSTED FOR 30.58 in
ALCO TARGET .077 18:11
BLANK TEST .000 18:11
INTERNAL STANDARD VERIFIED 18:12
ALCO TV 30.58 in .076 18:12
BLANK TEST .000 18:13
SUBJECT SAMPLE *BMB* .000 18:13
BLANK TEST .000 18:14
ALCO TV 30.59 in *2/8/11* .076 18:14
BLANK TEST .000 18:15

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

JANUARY 09, 2011
TIME 18:06

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.59 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~