VERIFICATION OF CALIBRATION REPORT

MAR 1 2009

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309

Supervisor/Operator Performing the Verification Procedure:

A

Name Shayne Calt ID# 4559 Date 3-7-09

Agency AST Phone # 733-2256

Instrument Location Mat-Su West

B

Alco S/N 75470 Target Value .076 High Pressure 700

Alco Test Value Average .077 (1st Alco) .077 (2nd Alco)

Signature

3/27/09

Under the Alaska Rules of Evidence, I certify that: (OVER)

(Do Not Write in this Area Below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Shayne M Calt, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

3/28/09

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MARCH 07, 2009 ✓

OPERATOR'S NAME: CMB 3/7/09
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATIONOFCAL
SUBJECT'S FIRST NAME/MI :
V
O.L. #: V
DEPT/AGENCY: TAL0
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .076 ✓
ALCO S/N: 75470 ✓

— BREATH ANALYSIS —

.076 ADJUSTED FOR 30.56 in

ALCO TARGET	.077	14:08
BLANK TEST	.000	14:09
INTERNAL STANDARD	VERIFIED	14:09
ALCO TV 30.56 in	.077 ✓	14:10
BLANK TEST	.000	14:10
SUBJECT SAMPLE CMB 3/7/09	.000	14:11
BLANK TEST	.000 ✓	14:12
ALCO TV 30.56 in	.077	14:12
BLANK TEST	.000	14:13

DIAGNOSTIC CHECK

700 psi

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MARCH 07, 2009 ✓
TIME 14:07

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	41c
BAROMETER:	30.56 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~