

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130309</u>
Name	<u>P. Nelson</u>	ID#	<u>4389</u> Date <u>5/10/10</u>
A	Agency	<u>AST</u>	Phone # <u>373-8337</u>
	Instrument Location	<u>MSW</u>	
Alco S/N		<u>31085</u>	Target Value <u>.075</u> High Pressure <u>700</u>
B	Alco Test Values	<u>.074</u>	<u>.074</u>
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<u>[Signature]</u>	
(OVER)			

*DMB*  
*5/24/10*

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10<sup>th</sup> day of June, 2010.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309  
COB 5/18/10

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

MAY 10, 2010

OPERATOR'S NAME: NELSON PATRICK  
OPERATOR'S NUMBER: 4389  
SUBJECT'S LAST NAME:  
VERIFICATION OF  
SUBJECT'S FIRST NAME/MI :  
CALIBRATION  
O.L. #: 123456  
DEPT/AGENCY: PLM0  
CASE/REPORT: 12-3456  
TEST TYPE: U  
ALCO TARGET VALUE: .075  
ALCO S/N: 31085

*BMB*  
*5/24/10*

### --- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.77 in		
ALCO TARGET	.074	16:05
BLANK TEST	.000	16:06
INTERNAL STANDARD	VERIFIED	16:06
ALCO TU 29.77 in	.074	16:07
BLANK TEST	.000	16:07
SUBJECT SAMPLE	.000	16:08
BLANK TEST	.000	16:09
ALCO TU 29.77 in	.074	16:09
BLANK TEST	.000	16:10

*BMB*  
*5/24/10*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

MAY 10, 2010  
TIME 16:01

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.77 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~