VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130 309

Supervisor/Operator Performing the Verification Procedure:

A

Name Ronny Simmons ID# 5123 Date 6-13-08

Agency AST Phone # 373-8324

Instrument Location Mat-Su West

B

Alco S/N X124609 Target Value .078 High Pressure 1200

Alco Test Value Average .081 (1st Alco) .081 (2nd Alco)

Signature RAS 6/27/08

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do Not Write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Ronny Simmons , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed

6/30/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 01/00 JUN 18 2008

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130309

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

JUNE 13, 2008 CDS 6/27/08

OPERATOR'S NAME:
SIMMONSRONNYA
OPERATOR'S NUMBER: 5123
SUBJECT'S LAST NAME:
VERIFICATIONOFCAL
SUBJECT'S FIRST NAME/MI :
D
O.L. #: 1
DEPT/AGENCY: PLM0
CASE/REPORT: 1
TEST TYPE: V
ALCO TARGET VALUE: .078
ALCO S/N: X124609

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.06 in

ALCO TARGET	.078	23:29
BLANK TEST	.000	23:30
INTERNAL STANDARD	VERIFIED	23:30
ALCO TV 30.06 in	.081	23:30
BLANK TEST	.000	23:31
SUBJECT SAMPLE	.000 CDS 6/27/08	23:31
BLANK TEST	.000	23:32
ALCO TV 30.06 in	.081	23:33
BLANK TEST	.000	23:33

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

JUNE 13, 2008
TIME 23:34

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.06 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~