


# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130309 ✓
Name	PATRICK NELSON	ID#	4389 Date 8-12-08 ✓
A Agency	AST	Phone #	373-8337
Instrument Location	MATSU WEST ✓		
B Alco S/N	X124609 ✓	Target Value	.078 ✓ High Pressure 650
Alco Test Values	.078 ✓ 1 <sup>st</sup> Alco	.077 ✓ 2 <sup>nd</sup> Alco	
Signature			 8/18/08


(OVER)

Under the Alaska Rules of Evidence, I certify that (type in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed.



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

8/20/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130309 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

AUGUST 12, 2008 ✓

OPERATOR'S NAME:

NELSON PATRICK

OPERATOR'S NUMBER: 4389

SUBJECT'S LAST NAME:

VERIFICATION OF

SUBJECT'S FIRST NAME/MI :

CALIBRATION

O.L. #: 123456

DEPT/AGENCY: PLND

CASE/REPORT: 12-34567

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X124609 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.46 in  
ALCO TARGET .076 16:26  
BLANK TEST .000 16:27  
INTERNAL STANDARD .078 ✓ 16:27  
ALCO TV 29.46 in .078 ✓ 16:27  
BLANK TEST .000 16:28  
SUBJECT SAMPLE .000 16:29  
BLANK TEST .000 16:29  
ALCO TV 29.46 in .077 ✓ 16:30  
BLANK TEST .000 16:31

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

AUGUST 12, 2008 ✓

TIME 16:24

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.46 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { } ~ ¡ ¢