

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130309 ✓

Name P. NELSON ID# 4389 ✓ Date 8/23/10 ✓

A Agency AST Phone # 746-9108

Instrument Location MSW

---

B Alco S/N X172971 ✓ Target Value .083 High Pressure 300

Alco Test Values .084 ✓ ~~0.083~~ PSND .084 ✓

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature [Signature]

(OVER)

BMB  
8/25/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>RD</sup> day of SEPT, 2010.

[Signature] (Notary Seal Stamp)

Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

AUGUST 23, 2010

OPERATOR'S NAME:  
NELSON PATRICK S  
OPERATOR'S NUMBER: 4389  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/MI :  
OF CALIBRATION  
O.L. #: 1234567  
DEPT/AGENCY: PL10  
CASE/REPORT: 12-34567  
TEST TYPE: V  
ALCO TARGET VALUE: .083  
ALCO S/N: X172971

*BMB*  
*8/25/10*

--- BREATH ANALYSIS ---

.083 ADJUSTED FOR 29.94 in  
ALCO TARGET .083 09:19  
BLANK TEST .000 09:21  
INTERVAL STANDARD VERIFIED 09:21  
ALCO TU 29.93 in .084 09:21  
BLANK TEST .000 09:22  
SUBJECT SAMPLE .080 09:22  
BLANK TEST .000 09:23  
ALCO TU 29.93 in .084 09:23  
BLANK TEST .000 09:24

*BMB*  
*8/25/10*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130309

AUGUST 23, 2010  
TIME 09:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.94 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢