

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

40808

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

FEBRUARY 27, 2011

OPERATOR'S NAME: *BMB*
CRABTREE/WILLIAM/A
OPERATOR'S NUMBER: 6316
SUBJECT'S LAST NAME: *4/4/11*
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: 123456
DEPT/AGENCY: B164
CASE/REPORT: VOC
TEST TYPE: T
ALCO TARGET VALUE: .007
ALCO S/N: X172958

--- BREATH ANALYSIS ---

.007 ADJUSTED FOR 29.51 in		
ALCO TARGET	.005	14:29
BLANK TEST	.000	14:30
INTERNAL STANDARD	VERIFIED	14:30
ALCO TV 29.51 in	.000	14:30
BLANK TEST	.000	14:31
SUBJECT SAMPLE	.000	14:32
BLANK TEST	.000	14:33
ALCO TV 29.51 in	.000	14:33
BLANK TEST	.000	14:34

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

FEBRUARY 27, 2011
TIME 14:19

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""'()*+,-./0123456789:;<=>@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~