

FEB 13 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

Supervisor/Operator Performing the Verification Procedure:

Name ERIC C. WALLA ID# 5505 Date 2008/02/10

A


Agency FORT GREELY POLICE DEPARTMENT Phone # 907-873-9135

Instrument Location 663 FIRST STREET FORT GREELY, ALASKA 99731

B

Alco S/N X301581 ✓ Target Value .080 ✓ High Pressure 600 PSI

Alco Test Values .075 .075  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature 

CSB  
2/21/08

(OVER)


Under the Alaska Rules of Evidence, I certify that:  
(Do Not write in the area below)

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Eric C Walla, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

02/22/08

Date



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

FEBRUARY 10, 2008 *CSB 2/11/08*

OPERATOR'S NAME:

WALLA ERIC C

OPERATOR'S NUMBER: 5505

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

FEBRUARY M

O.L. #: 20080210

DEPT/AGENCY: B164

CASE/REPORT: VOC10FEB2008

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: X301581

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 28.39 in		
ALCO TARGET	.075	09:47
BLANK TEST	.000	09:48
INTERNAL STANDARD	VERIFIED	09:48
ALCO TV 28.39 in	.075	09:48
BLANK TEST	.000	09:49
SUBJECT SAMPLE	.000 <i>CSB 2/11/08</i>	09:50
BLANK TEST	.000	09:51
ALCO TV 28.39 in	.075	09:51
BLANK TEST	.000	09:52

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY -  
DATAMASTER cdm 130404

FEBRUARY 10, 2008

TIME 09:44

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	28.39 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ J ' abcde fghi jklmno  
pqrstuvwxyz{ } ~