

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 16 2009

Supervisor/Operator Performing the Verification Procedure:	
Name	<u>William A. Crabtree</u> ID# <u>6316</u> Date <u>03/12/09</u>
Agency	<u>Fort Greely Police Department</u> Phone # <u>873-9134</u>
Instrument Location	<u>Bldg 663 Booking Room</u> ✓
DataMaster cdm S/N <u>130404</u> ✓	
Alco S/N	<u>X301572</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>1000</u>
Alco Test Values	<u>.079</u> ✓ 1 <sup>st</sup> Alco <u>.078</u> ✓ 2 <sup>nd</sup> Alco
Signature	<u>William A Crabtree</u> ✓
CSB 3/24/09	


(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, William Crabtree, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/24/09

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MARCH 12, 2009 ✓

OPERATOR'S NAME: *Spaulley*

CRABTREE/WILLIAM/A

OPERATOR'S NUMBER: 6316

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: L

DEPT/AGENCY: B164

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301572 ✓

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 28.74 in  
ALCO TARGET .075 01:04  
BLANK TEST .000 01:05  
INTERNAL STANDARD VERIFIED 01:05  
ALCO TV 28.74 in .079 ✓ 01:05  
BLANK TEST .000 01:06  
SUBJECT SAMPLE *COB 3P169* .000 01:07  
BLANK TEST .000 ✓ 01:08  
ALCO TV 28.74 in .078 ✓ 01:08  
BLANK TEST .000 01:09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MARCH 12, 2009 ✓

TIME 01:14

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 28.74 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~