

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAY 18 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130404</u>
Name	<u>BRAD REED</u>	ID# <u>6158</u> Date <u>5-15-09</u>
A Agency	<u>FORT Greely Police Dept.</u>	Phone # <u>873-9149</u>
Instrument Location	<u>Booking Room</u>	
B Alco S/N	<u>X 301572</u>	Target Value <u>.079</u> High Pressure <u>900</u>
Alco Test Values	<u>.079</u> 1 <sup>st</sup> Alco	<u>.080</u> 2 <sup>nd</sup> Alco
Signature	<u>[Signature]</u>	<u>003</u> <u>5/19/09</u>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27<sup>th</sup> day of May, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130404 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MAY 15, 2009 ✓

OPERATOR'S NAME:

REED BRAD G. *BRG*

OPERATOR'S NUMBER: 6158

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 00000000

DEPT/AGENCY: B164

CASE/REPORT: 00000000

TEST TYPE: T

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301572 ✓

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 28.87 in

ALCO TARGET	.076	10:32
BLANK TEST	.000	10:33
INTERNAL STANDARD	VERIFIED	10:33
ALCO TV 28.87 in	.079 ✓	10:34
BLANK TEST	.000	10:34
SUBJECT SAMPLE	.079 ✓	10:35
BLANK TEST	.000	10:36
ALCO TV 28.87 in	.079 ✓	10:36
BLANK TEST	.000	10:37

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

MAY 15, 2009 ✓

TIME 10:27

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 40c  
BREATH TUBE: 42c  
BAROMETER: 28.87 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#\$%^&\*~.,/0123456789:;<=>?@ABCDEFG  
HIJKLMNOPQRSTUVWXYZ\]^\_`abcde fghijklmnop  
qrstuvwxyz{|}~>